



Cadenza CX-C18

150 x 3 mm

HPLC COLUMN APPLICATION

Sibutramine (Anoretics)

シブトラミン (食欲抑制作用薬)



150 x 3 mm

A: water / formic acid = 100 / 0.1

B: water / acetonitrile / formic acid = 10 / 90 / 0.1

10-100%B (0-15min)

0.4mL/min (10MPa), 37°C, 260nm

5µL (1mg/mL, water / formic acid = 100 / 1)